

Recycling Separation of Oligosaccharide sample A

Instrument: Recycle Preparative HPLC, LC-9201

Column : JAIGEL-W252(20φ x 500 mm)

Flow rate : 3.5 mL/min

Eluent : Water (HPLC grade)

Sample volume : 200 mg/ 3 mL

Detector : RI (x 500 Range)

