

Recycling Separation of Glycans(Tagatose – Galactose)

Instrument: Recycle Preparative HPLC, LC-9201

Column : JAIGEL-W252 (20φ x 500 mm)

Flow rate : 3 mL/min

Eluent : Water

